

**Customer OEFA**

**Date of** recieve 26/04/22 Inspection 26/04/22  
**Model** 180 **Serial Number** 18A20138  
**Reason for service** 320C

**Case**
**Customer values**

Alarm N	0µg/m³	Alarm C	0p/L	Thr. Humidity	51%
Intervall	0 ... 1min	Location	1		
<b>Values before Service</b>		<b>pre-Comparison</b>		<b>Values after Service</b>	
Date of Inspection		Yes	No	Date of EK	
Run-Time	2766,9h	Ramp	<input type="checkbox"/>	Run-Time	2771,7h
Dryer-Time	2409,9h	Mass	<input type="checkbox"/>	Dryer-Time	2410,4h
Software Version		<b>Calibration</b>		Software Version	
7.80 E	DM180G 29.8.2013	Calibration	<input checked="" type="checkbox"/>	DC/v	70,0mV
DC/v	71,8mV	Ramp	<input type="checkbox"/>	DC_d	123,0mV
DC_d	124,8mV	Mass Tower	<input checked="" type="checkbox"/>	DC_h	134,0mV
DC_h	139,2mV	Check data	<input checked="" type="checkbox"/>	DC_diff	11,0mV
DC_diff	14,4mV	archiving	<input type="checkbox"/>	CO_h	0
CO_h	0			CO_d	0
CO_d	1			La_l	51mA
La_l	51mA			La_h	89mA
La_h	89mA			Flow	1,20L/min
Flow	1,20L/min			Imot	49,7%
Imot	52,0%			Vacuum	-50,0 kPa
Vacuum	-50,0 kPa			Flow Vacuum	4,70L/min
Flow Vacuum	4,70L/min			Voltage P-Valve	2.4 V
Voltage P-Valve	2.4 V			Date	ok
Date	ok			Time	ok
Time	ok				

**Service activities**

<b>Inspection &amp; Final Check</b>		<b>Modifications</b>	
Analog Inputs	<input type="checkbox"/> / <input type="checkbox"/>	Check Channel Lift	<input type="checkbox"/>
Keyboard	<input checked="" type="checkbox"/> / <input type="checkbox"/>	Yes <input type="checkbox"/> No <input checked="" type="checkbox"/>	<input type="checkbox"/>
Memorycard	<input checked="" type="checkbox"/> / <input type="checkbox"/>	Humidity TH:55%	<input type="checkbox"/>
Rinsing Air Contr.	<input checked="" type="checkbox"/>	Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	<input type="checkbox"/>
Pneumatic tightn.	<input checked="" type="checkbox"/> / <input type="checkbox"/>		<input type="checkbox"/>
O-Filter Check	<input checked="" type="checkbox"/> / <input type="checkbox"/>	7.80 E DM180G	<input type="checkbox"/>
Function Lift	<input checked="" type="checkbox"/> / <input type="checkbox"/>		<input type="checkbox"/>
Sensor Check	<input checked="" type="checkbox"/> / <input type="checkbox"/>		<input type="checkbox"/>
Fast-Mode OFF	<input checked="" type="checkbox"/>		<input type="checkbox"/>
Cleaning	<input type="checkbox"/> / <input type="checkbox"/>		<input type="checkbox"/>