

Customer OEFA

Date of **recieve** **Inspection**
26/04/22 26/04/22
Model 180 **Serial Number** 18A20143
Reason for service 319C

Case
Customer values

Alarm N	0µg/m³	Alarm C	0p/L	Thr. Humidity	51%
Intervall	0 ... 1min	Location	1		
Values before Service		pre-Comparison		Values after Service	
Date of Inspection		Yes	No	Date of EK	
Run-Time	2906,7h	Ramp	<input type="checkbox"/>	Run-Time	2913,0h
Dryer-Time	2727,3h	Mass	<input type="checkbox"/>	Dryer-Time	2727,3h
Software Version		Calibration		Software Version	
7.80 E DM180G 29.8.2013		Calibration	<input checked="" type="checkbox"/>	DC/v	27,6
DC/v	28,8mV	Ramp	<input type="checkbox"/>	DC_d	74,5mV
DC_d	75,6mV	Mass Tower	<input checked="" type="checkbox"/>	DC_h	90,0mV
DC_h	90,9mV	Check data	<input checked="" type="checkbox"/>	DC_diff	15,5mV
DC_diff	15,3mV	archiving	<input type="checkbox"/>	CO_h	0
CO_h	0			CO_d	0
CO_d	0			La_l	51mA
La_l	51mA			La_h	85mA
La_h	84mA			Flow	1,20L/min
Flow	1,20L/min			Imot	50,4%
Imot	50,4%			Vacuum	-50,0 kPa
Vacuum	-50,0 kPa			Flow Vacuum	4,70L/min
Flow Vacuum	4,70L/min			Voltage P-Valve	2.4 V
Voltage P-Valve	2.4 V			Date	ok
Date	ok			Time	ok
Time	ok				

Service activities

Inspection & Final Check		Modifications	
Analog Inputs	<input type="checkbox"/> / <input type="checkbox"/>	Check Channel Lift	<input type="checkbox"/>
Keyboard	<input checked="" type="checkbox"/> / <input type="checkbox"/>	Yes <input type="checkbox"/> No <input checked="" type="checkbox"/>	<input type="checkbox"/>
Memorycard	<input checked="" type="checkbox"/> / <input type="checkbox"/>	Humidity TH:55%	<input type="checkbox"/>
Rinsing Air Contr.	<input checked="" type="checkbox"/>	Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	<input type="checkbox"/>
Pneumatic tightn.	<input checked="" type="checkbox"/> / <input type="checkbox"/>		<input type="checkbox"/>
O-Filter Check	<input checked="" type="checkbox"/> / <input type="checkbox"/>	7.80 E DM180G	<input type="checkbox"/>
Function Lift	<input checked="" type="checkbox"/> / <input type="checkbox"/>		<input type="checkbox"/>
Sensor Check	<input checked="" type="checkbox"/> / <input type="checkbox"/>		<input type="checkbox"/>
Fast-Mode OFF	<input checked="" type="checkbox"/>		<input type="checkbox"/>
Cleaning	<input type="checkbox"/> / <input type="checkbox"/>		<input type="checkbox"/>