

Date&time: 23/04/2022 11:02

User name: Demo

Comment:

SDS Version: V4-1 Rev VII (07-08-2015)

Serial No.: 8HG20137

Model: 180D

Version: P

Firmware: 7.80

DC/v [mV]: 94.6

DC_d [mV]: 138.0

DC_h [mV]: 156.2

DC_diff [mV]: 18.2

CO_h [mV]: 0.0

CO_d [mV]: 0.0

LA_l [mA]: 51.0

LA_h [mA]: 93.0

GF:1.00

error code: 0

Im [%]: 42

weight [ug]: 0.0

volume [m3]: 0.001

temperature [°C]: 19.9

humidity [%]: 0.4

pressure [hPa]: 673.7

Informations:

"all status data of device are ok "

Field test is ok: 0.94 μm (accepted range: 0.9 μm to 1.1 μm)

