

Certificate of Analysis List

For request number 1449103

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
LZW9710.99 uS/cm	1104	N/A	N/A	Conductivity Standard Solution, 1413

Total Enclosures: 1

***Certificate of Analysis***

Page 1

COMMODITY: **Conductivity Standard Solution, 1413 uS/cm**COMMODITY NUMBER: **LZW9710.99** MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A1104****5/4/2021****5/4/2021**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
conductivity	1398 to 1426 uS/cm	1407.0 uS/cm

The expiration date is May 2022

The item LZW9710.99 is traceable to NIST standards SRM 999C Potassium Chloride
LOT N/A.

Certified by

Scott Als
Analytical Services Chemist