

# **Certificate of Analysis List**

For request number 1449092

<b>Catalog Number Entered</b>	<b>Lot Number Entered</b>	<b>Related Catalog Number</b>	<b>Related Lot Code</b>	<b>Description</b>
2283449	1085	N/A	N/A	Buffer Solution pH 4.01 ± 0.02

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Buffer Solution pH 4.01 ± 0.02**COMMODITY NUMBER: **2283449**

MANUFACTURE DATE:

LOT NUMBER: **A1085****4/6/2021**

DATE OF ANALYSIS:

**4/6/2021**

---

<i><b>TEST</b></i>	<i><b>SPECIFICATIONS</b></i>	<i><b>RESULTS</b></i>
pH of the solution @25C	3.985 to 4.025	4.0040

The expiration date is Mar 2025

The item 2283449 is traceable to NIST standards SRM 185i Potassium Hydrogen Phthalate LOT N/A.

Certified by

---

Scott Als  
Analytical Services Chemist