

Certificate of Analysis List

For request number 1502836

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2283449	0209	N/A	N/A	Buffer Solution pH 4.01 ± 0.02

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Buffer Solution pH 4.01 ± 0.02**COMMODITY NUMBER: **2283449**

MANUFACTURE DATE:

LOT NUMBER: **A0209****8/3/2020**

DATE OF ANALYSIS:

8/3/2020

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
pH of the solution @25C	3.985 to 4.025	3.9850

The expiration date is Jul 2024

The item 2283449 is traceable to NIST standards SRM 185i Potassium Hydrogen Phthalate LOT N/A.

Certified by

Scott Als
Analytical Services Chemist