

Certificate of Analysis List

For request number 1294709

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
1222149	0063	N/A	N/A	Buffer Solution Carbonate Type

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Buffer Solution Carbonate Type**COMMODITY NUMBER: **1222149**

MANUFACTURE DATE:

LOT NUMBER: **A0063****3/7/2020**

DATE OF ANALYSIS:

3/7/2020

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
pH of the solution	9.99 to 10.03	9.990

The expiration date is Aug 2022

The item 1222149 is traceable to NIST standards SRM 191d-1, 191d-11 Sodium Bicarbonate and Sodium Carbonate LOT N/A.

Certified by

Scott Als
Analytical Services Chemist