

Certificate of Analysis List

For request number 1463175

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
1222349	1117	N/A	N/A	Buffer Solution Phthalate Type

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Buffer Solution Phthalate Type**COMMODITY NUMBER: **1222349**

MANUFACTURE DATE:

LOT NUMBER: **A1117****5/5/2021**

DATE OF ANALYSIS:

5/5/2021

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
pH of the solution	3.98 to 4.02	4.000

The expiration date is Apr 2025

The item 1222349 is traceable to NIST standards SRM 185i Potassium Hydrogen Phthalate LOT N/A.

Certified by

Scott Als
Analytical Services Chemist