

# **Certificate of Analysis List**

For request number 1449251

<b>Catalog Number Entered</b>	<b>Lot Number Entered</b>	<b>Related Catalog Number</b>	<b>Related Lot Code</b>	<b>Description</b>
1222249	1123	N/A	N/A	Buffer Solution Phosphate Type

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Buffer Solution Phosphate Type**COMMODITY NUMBER: **1222249**

MANUFACTURE DATE:

LOT NUMBER: **A1123****5/10/2021**

DATE OF ANALYSIS:

**5/10/2021**

---

<i><b>TEST</b></i>	<i><b>SPECIFICATIONS</b></i>	<i><b>RESULTS</b></i>
pH of the solution	6.98 to 7.02	6.999

The expiration date is May 2024

The item 1222249 is traceable to NIST standards SRM 186-I-g, 186-II-g Potassium Dihydrogen Phosphate and Disodium Hydrogen Phosphate LOT N/A.

Certified by

---

Scott Als  
Analytical Services Chemist