

Certificate of Analysis List

For request number 1557673

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2283649	2056C	N/A	N/A	Buffer Solution pH 10.01 ± 0.02

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Buffer Solution pH 10.01 ± 0.02**COMMODITY NUMBER: **2283649**

MANUFACTURE DATE:

LOT NUMBER: **A2056C****3/2/2022**

DATE OF ANALYSIS:

3/2/2022

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
pH of the solution	9.995 to 10.035	10.0340

The expiration date is Mar 2023

The item 2283649 is traceable to NIST standards SRM 191d-1, 191d-11 Sodium Bicarbonate and Sodium Carbonate LOT N/A.

Certified by

Scott Als
Analytical Services Chemist