

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: Buffer Solution pH 4.01

COMMODITY NUMBER: 2283449

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A8313

11/14/2018

11/14/2018

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
pH of the solution @25C	3.985 to 4.025	3.9860

The expiration date is Nov 2022

The item 2283449 is traceable to NIST standards SRM 185i Potassium Hydrogen Phthalate LOT N/A.

Certified by

A handwritten signature in cursive script that reads "Scott Als".

Scott Als
Analytical Services Chemist