

# **Certificate of Analysis List**

For request number 1449100

<b>Catalog Number Entered</b>	<b>Lot Number Entered</b>	<b>Related Catalog Number</b>	<b>Related Lot Code</b>	<b>Description</b>
2283549	1124	N/A	N/A	Buffer Solution pH 7.00 ± 0.02

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **Buffer Solution pH 7.00 ± 0.02**COMMODITY NUMBER: **2283549**

MANUFACTURE DATE:

LOT NUMBER: **A1124****5/11/2021**

DATE OF ANALYSIS:

**5/11/2021**

---

<i><b>TEST</b></i>	<i><b>SPECIFICATIONS</b></i>	<i><b>RESULTS</b></i>
pH of the solution @ 25C	6.98 to 7.02	7.014

The expiration date is May 2023

The item 2283549 is traceable to NIST standards SRM 186-I-g, 186-II-g Potassium Dihydrogen Phosphate and Disodium Hydrogen Phosphate LOT N/A.

Certified by

---

Scott Als  
Analytical Services Chemist